

PRODUCT TECHNICAL DATASHEET

Product Profile & Instructions for Use

Warwick House, Heathcote Way, Warwick
CV34 6TE, United Kingdom
T: 0044 (0)1926 422427 www.sasco.co.uk
enquiries@sasco.co.uk

CMP Commode Pan / Bidet

Not a Medical Device

Dimensions : 359 x 338 x 130 mm
Volume : 3000 ml
Material : Green Polypropylene
Pack Size : 1 x 20



Intended Use

- Commode Pan and Bidet.
- Body waste disposal product.
- Reusable product.

Intended User

- Device to be used by patients, clinical, nursing, and care professionals.

Product Features

- Strong and durable.
- Guaranteed 1000 cycles.
- Can be cleaned in automatic washer disinfectors or hand washed.
- Smooth edges and surfaces.

Packaging Information


Catalogue Number	Case Bar Code (GTIN 14)	Primary Packaging	Secondary Packaging	Case Weight
CMP	15060098135291	Polythene bag 1 x 20	Cardboard case 1 x 20	8 kg

Product Symbols

Manufacturer	Catalogue number	Caution	Batch code	Use-by date	Non-sterile	Absence of natural rubber latex	Do not use if package is damaged	Keep dry	Keep away from sunlight

Additional Information

- Harmonised System – Product Tariff / Commodity Export Code 39.26.9097.
- Country of Origin United Kingdom.



PRODUCT TECHNICAL DATASHEET

Product Profile & Instructions for Use

Warwick House, Heathcote Way, Warwick
CV34 6TE, United Kingdom
T: 0044 (0)1926 422427 www.sasco.co.uk
enquiries@sasco.co.uk

Manufacturer



- **Warwick SASCo Ltd. Warwick House, Heathcote Way, CV34 6TE, United Kingdom.**
- **Telephone: 0044 (0)1926 422427 E-mail: enquiries@sasco.co.uk**

Distributor/Importer



Standards

- **Product complies with:** **BS 2588:1992 (Part 2) Portable sanitary pans for use in healthcare establishments.**
- **Company compliant to:** **BS EN ISO 17664-1:2021 Processing of health care products – Information to be provided by the medical device manufacturer.**
BS EN ISO 20417:2021 Medical devices. Information to be supplied by the manufacturer.
BS EN ISO 15223-1:2016 Medical Devices. Symbols to be used with medical device labels, labelling and information to be provided
BS EN ISO 13485:2016 Medical Devices. Quality management systems. Requirements for regulatory purposes.
BS EN ISO 14971:2019 Medical Devices. Application of risk management.
- **Referenced standards:** **BS EN ISO 15883 Washer disinfectors. Requirements.**

Instructions for Use

Storage and Handling

- **Do not drop.**
- **Store in original packaging.**
- **Store in clean and dry conditions.**
- **Store off the floor on a shelf or pallet.**
- **Shelf life 36 months.**

Preparation

- **Product supplied non-sterile.**
- **Inspect products before processing.**
- **Withdraw damaged products and report to manager.**
- **Rinse with warm water to remove any packaging residues.**
- **Product to be used with suitable commode chair.**

Cleaning and Disinfection

- **Position products individually, avoiding contact with other items.**
- **Position products to allow water penetration and drainage.**
- **Machine clean in washer disinfectors compliant with ISO 15883.**
- **Wash cycle at 80°C. Disinfection at 90°C.**

Document Ref: 3.231
Date: August 2021
Revision: 11/21-0



Product Profile & Instructions for Use

Warwick House, Heathcote Way, Warwick
CV34 6TE, United Kingdom
T: 0044 (0)1926 422427 www.sasco.co.uk
enquiries@sasco.co.uk

- Use approved detergents, enzymatic or biocide cleaners, suitable for cleaning medical devices.
- Use detergents and disinfectants with a range between 2pH and 13pH.
- Manual cleaning is permissible.
- Rinse with clean water.

Inspection Maintenance and Incident Reporting

- Visually inspect product to identify damage.
- Discard damaged product.
- Any serious incident that has occurred during the use of this product should be reported to the manufacturer.

Disposal

- End of life is determined by wear and tear, visual inspection, or damage to product.
- If product is contaminated treat as clinical waste and dispose of in accordance with clinical waste procedures.

Warnings



- Do not use abrasive powders or metal brushes as they may cause surface damage.
- Do not exceed 100°C.

Important Information for Processing

- The information provided by Warwick SASCo is the company's official instructions for this product.
 - It is the responsibility of the processor to ensure processing is performed in accordance with this information and that personnel, machines and equipment comply with relevant standards for processing.
 - Any deviation from this information requires evaluation for the effectiveness and potential adverse consequences.
-
- Additional information available on request contact: enquiries@sasco.co.uk

